



Republic of the Philippines  
 Department of Health  
**BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER**  
 Baguio City

**BIDS AND AWARDS COMMITTEE  
 SUPPLEMENTAL / BID BULLETIN**

Form No.: MCC-BAC-013

Bid Bulletin No.: 2020-008-III

Date: October 27, 2020

This Bid Bulletin is issued to modify or amend the Bid Documents of:

IB No. 2020-007-BAC III Procurement of Various Medical Equipment and Instruments

ORIGINAL	AMENDED
<b>Bid Security</b>  14.1. The Bid and bid security shall be valid until <b>February 11, 2021</b> . Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.	<b>Bid Security</b>  14.1. The Bid and bid security shall be valid until <b>March 3, 2021</b> . Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.
<b>HIGH-SPEED MICROMOTOR DRILL SYSTEM, 50,000RPM to 80,000RPM</b>  1-piece Electric Motor <ul style="list-style-type: none"> <li>• Motor Speed: 50,000RPM to 80,000RPM</li> </ul> <b>HIGH PIECE SURGICAL HANDPIECE</b> <ul style="list-style-type: none"> <li>• 2 pieces Highspeed surgery handpiece (straight and angled) for up to 50,000rpm to <b>80,000RPM</b> and 1:1 transmission, with spray nozzles and quick lock for instruments with a length from 70 mm</li> </ul>	<b>HIGH-SPEED MICROMOTOR DRILL SYSTEM, 50,000RPM to (90,000RPM)</b>  1-piece Electric Motor <ul style="list-style-type: none"> <li>• Motor Speed: 50,000RPM to <b>(90,000RPM)</b></li> </ul> <b>HIGH PIECE SURGICAL HANDPIECE</b> <ul style="list-style-type: none"> <li>• 2 pieces Highspeed surgery handpiece (straight and angled) for up to 50,000rpm to <b>(90,000RPM)</b> and 1:1 transmission, with spray nozzles and quick lock for instruments with a length from 70 mm</li> </ul>
<b>AUDITORY BRAIN STEM RESPONSE ABR</b>	<b>AUDITORY BRAINSTEM RESPONSE ABR</b>
<b>BLOOD PRESSURE APPARATUS</b>  <b>Stand model Blood Pressure</b> <ul style="list-style-type: none"> <li>• 150mm dial diameter</li> <li>• Equipped with PU/Cloth-Velcro cuff for adults with plug-in adapter for quicker cuff exchange</li> </ul>	<b>BLOOD PRESSURE APPARATUS</b>  <b>Stand model Blood Pressure</b> <ul style="list-style-type: none"> <li>• 150mm dial diameter (<b>range: 145mm – 155mm</b>)</li> <li>• Equipped with PU/Cloth-Velcro cuff for adults with plug-in adapter for quicker cuff exchange</li> </ul> <b>Warranty: One (1) year</b>
<b>All other specifications remain unchanged.</b>	

**Addendum**

Additional Forms:

Statement of All On-Going Contract

Statement of Single Largest Completed Contract Similar to the Contract to be Bid.

OTHERS:

Questions:

1 Set Suction Pump- 60 liters is the speed, not the capacity.

Lot 6 Aerosol Evacuation System- Is the filter change alarm required?

- Yes, it is.

**BIDS AND AWARDS COMMITTEE**

**SGD. MARIA LORENA L. SANTOS, MD**

Chairperson

**SGD. ARNOLD WANSON, MD**

Vice Chairperson

**SGD. ROMMEL C. VERGANIO**

Member

**SGD. RICHARD ARCEO, MD**

Member

**SGD. MIGUELA FONGAFONG**

Member

**Statement of Single Largest Completed Contract (SLCC) Similar to the Contract to be Bid**

This is to certify that \_\_\_\_\_ has the following completed contract 3 years prior to the date of submission and receipt of bids.

Date of Contract	Contracting Party	Name of Contract	Kind of goods sold	Amount of Contract	Date of delivery / End User's Acceptance	Date of Official Receipt	Bidder is A. Manufacturer B. Supplier C. Distributor

\_\_\_\_\_  
Signature over Printed Name of Authorized Representative

Date \_\_\_\_\_

**Statement of: (i) On-going contracts and (ii) Awarded but not yet Started Contracts**

This is to certify that \_\_\_\_\_ has the following ongoing and awarded but not yet started contracts.

Date of Contract	Contracting Party	Name of Contract	Kind of goods sold	Amount of Contract	Value of outstanding Contracts	Bidder is A. Manufacturer B. Supplier C. Distributor

Signature over Printed Name of Authorized Representative

Date \_\_\_\_\_