

	Republic of the Philippines Department of Health BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER Baguio City	
	BIDS AND AWARDS COMMITTEE SUPPLEMENTAL/BID BULLETIN	
	Form No.: MCC-BAC-013 Revision No.: 1 Effectivity Date: September 1, 2016	
Bid Bulletin No.: 2020-005 Date: November 24, 2020		
<p>This Bid Bulletin No. 2020-005 is issued to modify or amend the Bid Documents of the project Supply and Delivery of Various Medical Equipment under Bayanihan Act 2 This shall form integral part of said Bid Documents.</p>		
ORIGINAL SPECS	AMENDED	
Supply and Delivery of One (1) unit Brand New Portable X-ray Machines (Retrofitted from Analog to Digital) with Flat Panel Detector Flat panel detector: wireless and cassette size <input type="checkbox"/> Scintillator: casium iodide (Csi) Portable work station <input type="checkbox"/> At least Core i5 processing unit	Supply and Delivery of One (1) unit Brand New Portable X-ray Machines (Retrofitted from Analog to Digital) with Flat Panel Detector Flat panel detector: wireless and cassette size <input type="checkbox"/> Scintillator: cesium iodide (Csi) Portable work station <input type="checkbox"/> At least 10th Generation, 6 cores, 12 threads and 12 mb Cache processing unit	
Supply and Delivery of Ten (10) units Movable Negative Pressure Machine with ULPA or Ultra Low Particulate Filter Accessories: 2 - 3meters aluminum exhaust tube	Supply and Delivery of Ten (10) units Movable Negative Pressure Machine with ULPA or Ultra Low Particulate Filter Accessories: 2 - 3meters aluminum exhaust tube or centrifugal fan for Negative Pressure	

BIDS AND AWARDS COMMITTEE I

(SGD)LARRY C. HAYA, MD
 Chairperson

(SGD)JERILYN L. DULAY, MD
 Vice Chairperson

(SGD)BERNADETTE B. VALDEZ, MD
 Member

(SGD)DENNIS M. DOMINGUEZ, MD
 Member

(SGD)ROMMEL PAUL G. FLORES
 Member

(SGD)JOYCE H. SORIANO
 Member

(SGD)ANDREW S. SIB-ATEN
 Member

Approved by:

(SGD)RICARDO B. RUNEZ, JR., MD, FPCS, MHA, CSEE
 Medical Center Chief II