

	Republic of the Philippines Department of Health <b>BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER</b> Baguio City	
	<b>BIDS AND AWARDS COMMITTEE          SUPPLEMENTAL/BID BULLETIN</b>	
	Form No.: MCC-BAC-013 Revision No.: 1 Effectivity Date: September 1, 2016	
<b>Bid Bulletin No.: 2021-007</b> <b>Date: August 31, 2021</b>		
<p>This Bid Bulletin No. 2021-007 is issued to modify or amend the Invitation to Bid of the projects:  <b>1.) One (1) Lot Supply, Installation and Commissioning of Gas Booster and Refilling Station</b></p> <p>This shall form integral part of said Invitation to Bid.</p>		
<b>TECHNICAL SPECIFICATIONS</b>		
<b>ORIGINAL SPECS</b>	<b>AMENDED</b>	
ONE (1) PACKAGE GAS BOOSTER – OXYGEN MEDIUM SPECIFICATIONS: - Flow Capacity: 187-283 Lpm	ONE (1) PACKAGE GAS BOOSTER – OXYGEN MEDIUM SPECIFICATIONS: - Flow Capacity: <b>at least</b> 283 Lpm	
ONE (1) LOT CENTRALIZED CONTROLLER (1 LOT)  Details: - Ten (10) inches Touch Screen, HMI Display  - Automatic Operation of Rix Booster - Maintenance Running hours of Rix Booster	ONE (1) LOT CENTRALIZED CONTROLLER <b>(SEPARATED OR INCORPORATED IN THE GAS BOOSTER)</b> Details: - Ten (10) inches Touch Screen, HMI Display or <b>manual operation</b> - Automatic Operation of <b>Rix</b> Booster - Maintenance Running hours of <b>Rix</b> Booster	
<b>CLARIFICATIONS</b>		
<b>QUESTIONS</b>	<b>ANSWER</b>	
What if the package is from a G7 Country but the plant is manufactured in China?	- The supplier should provide certification from a G7 country	
<b>INCLUSIONS</b>		
Please find attached drawings for the project.		

**BIDS AND AWARDS COMMITTEE I**

**SGD**  
**LARRY C. HAYA, MD**  
 Chairperson

**SGD**  
**JERILYN L. DULAY, MD**  
 Vice Chairperson

**SGD**  
**VALORIE B. PRESCO, RN**  
 Member

**SGD**  
**DENNIS M. DOMINGUEZ, MD**  
 Member

**SGD**  
**JOSELITO M. DATUD, RN**  
 Member

**SGD**  
**RANDY B. SAMONTE**  
 Member

**SGD**  
**ANDREW S. SIB-ATEN**  
 Member

Approved by:

**SGD**  
**RICARDO B. RUNEZ, JR., MD, FPCS, MHA, CSEE**  
 Medical Center Chief II