



Republic of the Philippines
Department of Health
BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER
Baguio City

**BIDS AND AWARDS COMMITTEE
SUPPLEMENTAL/BID BULLETIN**

Form No.: MCC-BAC-013
Revision No.: 1
Effectivity Date: September 1, 2016

Bid Bulletin No.: 2022-001
Date: January 17, 2022

This Bid Bulletin No. 2022-001 is issued to modify or amend the Bid Documents of the project:

1.) Supply, Delivery, Installation and Commissioning of One (1) unit Brand New Multi-Slice CT Scan Machine (Minimum of 16 Slice) & One (1) unit Ultrasound Machine (IB No. 2022-001 BAC I)

This shall form integral part of said Bid Documents.

NAME OF PROJECT

ORIGINAL SPECS	AMENDED
Supply, Delivery, Installation and Commissioning of One (1) unit Brand New 16 Slice CT Scan and One (1) unit Ultrasound Machine (IB No. 2022-001 BAC I)	Supply, Delivery, Installation and Commissioning of One (1) unit Brand New Multi-Slice CT Scan Machine (Minimum of 16 Slice) and One (1) unit Ultrasound Machine (IB No. 2022-001 BAC I)

SCHEDULE OF REQUIREMENTS

ORIGINAL SPECS	AMENDED
Delivery Period: Ninety (90) Calendar days upon Receipt of Notice to Proceed	Delivery Period: One Hundred Twenty (120) Calendar days upon Receipt of Notice to Proceed

TECHNICAL SPECIFICATIONS

ORIGINAL SPECS	AMENDED
BRAND NEW 16 SLICE CT SCAN MACHINE	BRAND NEW-MULTI-SLICE CT SCAN MACHINE (MINIMUM OF 16 SLICE)
2. Detector Detector material: Gadolinium OS solid	2. Detector Detector material: Gadolinium OS solid or its equivalent
4. X-ray Tube ... Anode Heat heat dissipation: at least 5.3MHU	4. X-ray Tube ... Anode Heat Capacity : at least 5.3MHU
7. Reconstruction and Console ... Image Recon Time: Up to 20 images/second w/out iterative	7. Reconstruction and Console ... Image Recon Time: at least 20 images/second w/out iterative
7.1 Computer (Console) ... Autofilm	7.1 Computer (Console) ... Autofilm or its equivalent
9. Clinical Enhancements and Applications 3-D SSD MIP/MinIP, Smart Planning, MPR, Insert MPR, Curved MPR ,AutoFilm AutoVoice ,CT angiography, Volume rendering, DoseRight Pediatric protocols, Virtual endoscopy, iPlanning, iBatch LNA: Lung nodule, CTC: Iterative Recon, VA (Vessel analysis), Bolus tracking, SAS, MPPS Worklist	9. Clinical Enhancements and Applications 3-D SSD MIP/MinIP, Smart Planning, MPR, Insert MPR, Curved MPR ,AutoFilm AutoVoice ,CT angiography, Volume rendering, DoseRight or its equivalent Pediatric protocols, Virtual endoscopy, iPlanning, iBatch or its equivalent LNA: Lung nodule, CTC: Iterative Recon, VA (Vessel analysis), Bolus tracking, SAS, MPPS Worklist or its equivalent
10. Other Requirements Dual Barrel Power Injector Two (2) years warranty on parts and services	10. Other Requirements Single or Dual Barrel Power Injector Minimum of Two (2) years warranty on parts and services

BIDS AND AWARDS COMMITTEE I

SGD

LARRY C. HAYA JR, MD

Chairperson

SGD

JERILYN L. DULAY, MD

Vice Chairperson

SGD

VALORIE B. PRESCO, RN

Member

SGD

DENNIS M. DOMINGUEZ, MD

Member

SGD

JOSELITO M. DATUD, RN

Member

SGD

RANDY B. SAMONTE

Member

SGD

ANDREW S. SIB-ATEN

Member

Approved by:

SGD

RICARDO B. RUNEZ, JR., MD, FPCS, MHA, CESE

Medical Center Chief II