



Republic of the Philippines
Department of Health
BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER
Baguio City

BIDS AND AWARDS COMMITTEE
SUPPLEMENTAL / BID BULLETIN

Form No.: MCC-BAC-013

Bid Bulletin No.: **2022-018 BAC III**

Date: May 26, 2022

This Bid Bulletin is issued to modify or amend the Bid Documents of:

PROCUREMENT OF ONE(1) UNIT CARTRIDGE-BASED NUCLEIC-ACID
AMPLIFICATION TEST/ REAL- TIME PCR EQUIPMENT FOR DETECTING
SARSCOV-2 , IB 2022-020 BAC III

ORIGINAL	AMENDED
<u>Schedule of Requirements</u>	
-30 Calendar Days upon receipt of NTP.	-60 Calendar Days upon receipt of NTP.
<u>Technical Specifications</u>	
2. Usage endorsed globally by WHO, and evaluated and validated by RITM (please provide documents)	2. Usage endorsed globally by WHO, and evaluated and validated by RITM (<u>Kits</u>) (please provide documents)
3. Real-time PCR technology with 6-channel excitation and detection optics integrating sample preparation, amplification, and detection	3. Real-time PCR technology with <u>6-10</u> channel excitation and detection optics integrating sample preparation, amplification, and detection
5. Capable for adaption and use as multi-testing for other diseases (primarily, HIV, TB, MDR-TB and HCV Viral Load tests prequalified by WHO) in integrated laboratory network	5. Capable for adaption and use as multi-testing for other diseases (primarily, <u>HIV Viral Load, HCV Viral Load and SARS-COV2</u> tests prequalified by WHO) in integrated laboratory network
8. With latest Computer system or latest Laptop interface	8. With latest Computer system or latest Laptop interface (<u>connectivity to the machine</u>)
13. Provide UPS and AVR (3KVA)	13. Provide <u>UPS (1KVA) and 3 prong AVR (3KVA)</u>
15. Free calibration kit to be performed 1-year after installation.	15. Free calibration kit to be performed <u>2-years, yearly calibration.</u>
	19. 24/7 physical / virtual availability of service engineer for troubleshooting incidents.
	20. To provide loaner unit in case of machine breakdown.

	21. ISO certification of the Manufacturer
	22. Calibration Certificate and Sticker upon delivery
	23. One (1) Starter pack (10 kits).

BIDS AND AWARDS COMMITTEE III

(SGD)

ARMEN R.POSADAS, MD

Chairperson

(SGD)

ALLAN U. FACTORA, RN

Vice-chairperson

(SGD)

CARLOS F. DURLAO II, MD

Member

(SGD)

MAUREEN P. CUALAN

Member

(SGD)

MIGUELA E. FONGAFONG, RN

Member

Approved by:

(SGD)

RICARDO B. RENEZ, JR., MD, FPCS, MHA, CESE

Medical Center Chief II