



Republic of the Philippines  
 Department of Health  
**BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER**  
 Baguio City

**BIDS AND AWARDS COMMITTEE  
 SUPPLEMENTAL / BID BULLETIN**

Form No.: MCC-BAC-013

Bid Bulletin No.: **2022-019 BAC III**

Date: May 26, 2022

This Bid Bulletin is issued to modify or amend the Bid Documents of:

**PROCUREMENT OF ONE(1) UNIT UPRIGHT AUTOCLAVE, IB 2022-021 BAC III**

ORIGINAL	AMENDED
<b><u>Technical Specifications</u></b>	
7. Operating Range: 105°C to <b>135°C</b>	7. Operating Range: 105°C to <b><u>135 °C -138°C</u></b>
8. Display Range: 5° C to <b>137 °C</b>	8. Display Range: 5° C to <b><u>135 °C -138 °C</u></b>
12. Maximum Allowable: <b>38 psi (0.26 MPa)</b>	12. Maximum Allowable: <b><u>38psi-40psi (0.26-0.28MPa)</u></b>
14. Overall (W x D x H): <b>20.1" x 20.8" x 40.9" to 22" x 23" x 42"</b>	14. Overall (W x D x H): <b><u>20.1"-22" x 20.8"-25"x40.9"-50"</u></b>
16. Dual-sensing lid interlock, over-pressure power cutoff, over-temperature power cutoff, over-temperature power cutoff, low-water power cutoff, mechanical pressure relief valve, lid-closure detector, over-current detector and circuit breaker.	16. Dual-sensing lid interlock, over-pressure power cutoff, over-temperature power cutoff, over-temperature power cutoff, low-water power cutoff, mechanical pressure relief valve, lid-closure detector, over-current detector and circuit breaker <b><u>or its equivalent</u></b>
18. <b>Stainless Steel</b> : 2 pcs Mesh wire baskets, 2 pcs buckets (pails), and 2 pcs round test tube racks	18. <b><u>Medical Grade Stainless Steel (316L)</u></b> : 2 pcs Mesh wire baskets, 2 pcs buckets (pails), and 2 pcs round test tube racks
20. Source: single phase ( <b>220V</b> )	20. Source: single phase ( <b><u>220V-240V</u></b> )
21. Consumption: <b>2000 watts (2000 VA)</b>	21. Consumption: <b><u>2000watts-6000watts (2000 VA-6000 VA)</u></b>
22. Provision of 3KVa 3 prong socket with grounding plug Automatic Voltage Regulator	22. Provision of 3KVa 3 prong socket with grounding plug Automatic Voltage Regulator with <b><u>circuit breaker</u></b>
	<b><u>27. ISO certification of the Manufacturer</u></b>
	<b><u>28. Calibration Certificate and Sticker upon delivery</u></b>

**BIDS AND AWARDS COMMITTEE III**

**ARMEN R.POSADAS, MD**  
Chairperson

**ALLAN U. FACTORA, RN**  
Vice-chairperson

**CARLOS F. DUMLAO II, MD**  
Member

**MAUREEN P. CUALAN**  
Member

**MIGUELA E. FONGAFONG,RN**  
Member

**Approved by:**

**RICARDO B. RUNEZ, JR., MD, FPCS, MHA, CESE**  
Medical Center Chief II