

	Republic of the Philippines Department of Health BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER Baguio City	
	BIDS AND AWARDS COMMITTEE SUPPLEMENTAL / BID BULLETIN	Form No.: MCC-BAC-013

Bid Bulletin No.: **2023-002-BAC III**

Date: February 8, 2023

This Bid Bulletin is issued to modify or amend the Bid Documents of:

PROCUREMENT OF ONE (1) SET BODY PLETHYSMOGRAPHY (BODY BOX)

IB 2023-002 BAC III

ORIGINAL	AMENDED
Section VII. SCHEDULE OF REQUIREMENTS	
30 Calendar Days upon receipt of Notice to Proceed.	90 Calendar Days upon receipt of Notice to Proceed.
Section VII. TECHNICAL SPECIFICATIONS	
2. SPIROMETRY	
b. Capable of different flowmeter configurations (pneumotach or turbine).	b. Capable of different flowmeter configurations (pneumotach or turbine or <u>ultrasonic</u>).
e. Capable to download raw data of 6 m w t thru a software app convertible to pdf	e. Capable to download raw data of <u>6-MWT (6 minute walk test)</u> thru a software app convertible to pdf
6. GAS ANALYZERS	
d. O2: Paramagnetic	d. O2: Paramagnetic <u>and/or electrochemical</u>
7. HARDWARE	
a. 17x34x42 cm or smaller	a. <u>Dimension: Manufacturer's specification</u>
b. UPS – 1.2kva	c. UPS – <u>at least</u> 1.2kva
8. POWER REQUIREMENT	
a. 100-240V, 50/60 Hz	a. 100-240V, 50/60 Hz (<u>Auto Volt</u>)
9. ENVIRONMENTAL SENSORS	
a. Temperature: 0°C to +50°C	a. Temperature: <u>within the range of</u> 0°C to +50°C
c. Humidity: 0-100%	c. Humidity: <u>within the range of</u> 0-100%
10. PC REQUIREMENT	
a. Monitor: 36 inch or more , with resolution 1920x1080 pixel Full HD, 2 or more USB ports.	b. Monitor: 36-inch or <u>24-inch and to provide 36-inch(or bigger) external slave monitor</u> with resolution 1920x1080 pixel Full HD, 2or more USB ports.

13. BODY BOX PHYSICAL FEATURES	
a. Dimension: Not less than (86x 185 x 77)cm, (WxHxD)	a. Dimension: Not less than (86x 170 x 71)cm, (WxHxD)
b. Volume: not less than 1500L	b. Volume: within 850 -1500L
c. With 3 dimension swivel arm (Flexible and height-adjustable)	h. With Adjustable and flexible swivel arm (Flexible and height-adjustable)
13. OTHERS	
	<u>i. Updated/valid FDA Certificates (CPR/CMDN and CMDR whichever is applicable) to be required during post-qualifications</u>

BIDS AND AWARDS COMMITTEE III

SGD
ARMEN R.POSADAS, MD
Chairperson

SGD
ALLAN U. FACTORA, RN
Vice-chairperson

SGD
MAUREEN P. CUALAN, MPA
Member

SGD
DENISH MITZ C. TALAVERA, MD
Member

SGD
STACEY DENISE D. ALZATE,MD
Member

SGD
ENDRIK H. SY, MD
Member

SGD
CORA L. DIAZ, RMT
Member

Approved by:

SGD
RICARDO B. RUNEZ, JR., MD, FPCS, MHA, CESE
Medical Center Chief II