

	Republic of the Philippines Department of Health BAGUIO GENERAL HOSPITAL AND MEDICAL CENTER Baguio City	
	BIDS AND AWARDS COMMITTEE SUPPLEMENTAL / BID BULLETIN	Form No.: MCC-BAC-013

Bid Bulletin No.: **2023-014 BAC III**

Date: June 20, 2023

This Bid Bulletin is issued to modify or amend the Bid Documents of:

**PROCUREMENT OF ONE (1) LOT CARDIOPULMONARY EXERCISE
TEST (CPET) MACHINE, RE-BID**

IB 2023-016 BAC III

ORIGINAL	AMENDED
Section I. Invitation to Bid	
9. Bid opening shall be on <i>June 27, 2023, 9:30 a.m. at the Small Conference Room, 2nd floor Cancer Center, BGHMC</i> , Baguio City and/or via WebEx Meeting. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.	9. Bid opening shall be on <i>June 27, 2023, <u>9:30 a.m. at the Small Conference Room, 2nd floor Cancer Center, BGHMC</u></i> , Baguio City and/or via WebEx Meeting. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
Section VI. Schedule of Requirements	
30 -Calendar Days from receipt of Notice to Proceed.	<u>90</u> Calendar Days from receipt of Notice to Proceed.
Section VII. Technical Specifications	
A. TECHNOLOGY	
5. Flow range: 0-18 l/s	5. Flow range: <u>within</u> 0-18 l/s
6. Accuracy: ±3%	6. Accuracy: <u>±2%</u>
E. THREADMILL FOR CPET	
4.1. 0.1 to 15 mph or 2.0 to 24 kph	4.1. <u>Within</u> 0.1 to 15 mph or 2.0 to 24 kph
9.1. Complied with class S, I according to at least ISO 20957-1 standards	9.1. Complied with class S, I according to at least ISO 20957-1 standards <u>or EN 957-1</u>

F. CYCLE ERGOMETER	
1.1. Minimum load: at least 7 watts	1.1. Minimum load: at least 6 watts

BIDS AND AWARDS COMMITTEE III

SGD

ARMEN R.POSADAS, MD

Chairperson

SGD

ALLAN U. FACTORA, RN

Vice-chairperson

SGD

MAUREEN P. CUALAN, MPA

Member

SGD

DENISH MITZ C. TALAVERA, MD

Member

SGD

LEANNE C. ACOSTA, MD

Member

SGD

ENDRIK H. SY, MD

Member

SGD

CORA L. DIAZ, RMT

Member

Approved by:

SGD

RICARDO B. RUNEZ, JR., MD, FPCS, MHA, CESE

Medical Center Chief II